clinell®

Touch-free Hand Disinfection




Set up guide

Carton contents

- 1. Box containing Clinell dispenser unit parts (including dispenser key)
- 2. Box containing backboard
- 3. 4 x D batteries
- 4. Drip tray (optional use)



Backboard



Dispenser


Batteries

Step one

Placement of the backboard

Place backboard against wall in desired location

- 1. On the wall, mark through four drill holes A, B, C & D with a pencil.
- 2. Drill holes into the wall according to four marked points.
- 3. Insert 4 x rawl plugs into drilled holes.
- **4.** Remove adhesive backing from backboard and carefully line up holes A, B, C & D with drill holes.
- 5. Push backboard firmly against wall for 30 seconds*.
- 6. Gently peel away protective adhesive on front of backboard.
- 7. Dispose of all packaging as per local regulations.





^{*}Caution: Backboard adhesive may damage walls on removal.

Step two


Mounting the dispenser onto the backboard

- 1. Loosely screw 2 x screws through backboard into holes marked A & B.
- 2. Ensure battery cover is removed from dispenser unit.
- 3. Unpeel adhesive film from back of dispenser unit.
- **4.** Carefully hook the open dispenser unit onto screws in holes A & B, ensuring unit doesn't stick to backboard.
- **5.** Line up holes C & D with holes in backboard and press unit firmly in position for 30 seconds.
- 6. Screw 2 x screws through holes C & D into backboard, tighten screws in holes A & B.













Step three

Dispenser set-up

- 1. Remove cap from gel refill cartridge by pulling the cap firmly.
- Turn cartridge upside down, ensuring label is facing towards the back of the dispenser unit and click into place.
- 3. Insert 4 x D batteries and replace battery cover.
- 4. Close dispenser unit and click shut.
- **5.** Line up drip tray clips and press in the left side, followed by the right side until firmly locked into position.
- **6.** Place hand under dispenser unit multiple times to activate gel (green light signifies unit is in operation).















Top tip

Place your Clinell Touch-free Hand Disinfection wall mounted dispenser near entrances and coffee/dining areas for maximum compliance.

For further information visit:

www.gamahealthcare.com/hand-disinfection

ub your fingertips
to opposing palm

ub each thumb
opposing hand

regul

ub your h alm to pa

